 Department of Microbiology Quality Manual	Policy # Mi_SER_EXTQC	Page 1 of 5
	Version: 1.1 CURRENT	
Section: Serology Procedures	Subject Title: Serology External Control Selection Criteria Revision Date: 1/15/2024 Next Review Date: 11/15/2025	
Prepared by QA Committee		
Issued by: Laboratory Manager		
Approved by Laboratory Director: Microbiologist-in-Chief		

Uncontrolled When Printed



TABLE OF CONTENTS

Introduction..... 2

Procedure..... 2

Related Documents 3

Record of Edited Revisions 4

  Department of Microbiology Quality Manual	Policy # Mi_SER_EXTQC	Page 2 of 5
	Version: 1.1 CURRENT	
Section: Serology Procedures	Subject Title: Serology External Control Selection Criteria	

Introduction

External Controls are samples external to tests kit provided with known expected results to assess the accuracy of a test. External controls are run according to individual test procedure but often tested on new test kit lot numbers, shipments or after major maintenance or deviation from standard conditions occurs.

Procedure

Use the chart below to determine which serology external controls to run by analyte.

**Class must be specified when order is placed.

Select Packaging	Product	Class	Analyte, Test Method
VIROTROL I Cat # 00101	1 x 5ml 10 x 4 ml	F	HBsAg, Abbott Architect Anti-HIV-1, Abbott Architect Anti-HTLV-I, Abbott Architect Anti-HBcore, Abbott Architect Anti-HCV, Abbott Architect Anti-CMV Total EIA, Immucor
VIROTROL II Cat # 00104	1 x 5 ml	B+	Anti-HBs, Abbott Architect Anti-HAV, Abbott Architec
VIROTROL III	1 x 5 ml	B	Anti-HBc IgM, Abbott Architect Anti-HAV IgM, Abbott Architect
VIROTROL HIV-2 Cat # 00105C	1 x 5 ml	C	HIV-2 Abbott Architect
VIROTROL HIV-1 Ag Cat # 00108A	1 x 5 ml	A	HIV-1 (P24) Ag Abbott Architect

UNIVERSITY HEALTH NETWORK/MOUNT SINAI HOSPITAL, DEPARTMENT OF MICROBIOLOGY

*NOTE: This document is **Uncontrolled When Printed**.*

Any documents appearing in paper form that do not state "CONTROLLED COPY" in red print are not controlled and should be checked against the document (titled as above) on the server prior to use.

Management System\UHN_Mount Sinai Hospital Microbiology\Standard Operating Procedures\Serology Procedures\

VIROTROL SYPHILIS Cat# 00071X	1 x 5 ml	N/A	Syphilis TP Ab Abbott Architect
VIROCLEAR Cat # 00112	1 0x 4 ml	N/A	All(except Rubella IgG)
VIROTORCH Cat# 000109	1 x 5 ml \$ 300.00	N/A	CMV IgG Abbott Architect
VIROTORCH-M Cat# 000117	1 x 5 ml \$ 438.00	N/A	CMV IgM Abbott Architect

Related Documents



Department of Microbiology

Quality Manual

Policy # Mi_SER_EXTQC

Page 5 of 5

Version: 1.1 CURRENT

Section: Serology Procedures

Subject Title: Serology External Control Selection Criteria

UNIVERSITY HEALTH NETWORK/MOUNT SINAI HOSPITAL, DEPARTMENT OF MICROBIOLOGY

*NOTE: This document is **Uncontrolled When Printed**.*

Any documents appearing in paper form that do not state "CONTROLLED COPY" in red print are not controlled and should be checked against the document (titled as above) on the server prior to use.

Management System\UHN_Mount Sinai Hospital Microbiology\Standard Operating Procedures\Serology Procedures\